

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1306

Name: RITTER 1  
Sterilizer ID: V2178198  
10/27/2021 09:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.1 F 0.0  
2:00 183.3 F 2.9  
4:00 185.0 F 7.1  
6:00 207.5 F 12.5  
8:00 223.9 F 7.4  
10:00 235.1 F 13.1  
12:00 237.9 F 10.5  
14:00 235.4 F 8.6  
16:00 250.2 F 15.7  
18:00 262.4 F 22.7  
20:00 272.2 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.0 F 28.9  
2:00 272.6 F 29.4  
3:00 271.8 F 29.3  
4:00 271.9 F 29.0

Min 271.4 F 28.8  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.0 F 0.0  
5:00 239.1 F 0.0  
10:00 242.2 F 0.0  
15:00 231.9 F 0.0  
20:00 180.0 F 0.0  
25:00 164.3 F 0.0  
Dry Cycle Aborted

FILLING: 1:29  
HEATING: 20:11  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 26:17  
TOTAL CYCLE: 00:55:01

10/27/2021 10:08 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01