
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001679
12:13 10/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:38
STERILIZING 3:42
136.4 C 223kPa 3:42
136.2 C 223kPa 4:12
136.1 C 222kPa 4:42
136.4 C 224kPa 5:12
136.3 C 225kPa 5:42
136.3 C 224kPa 6:12
136.3 C 224kPa 6:42
136.4 C 223kPa 7:12
136.4 C 223kPa 7:12
VENTING 7:12
AIR DRYING 8:02
CYCLE COMPLETE 8:05

Digital Signature #
7D4C03CB16475972
