
Midmark M11 - v1.0.5
Total Cycles: 3537

Name: RITTER2
Sterilizer ID: V2313228
04/17/2024 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.9 F 0.0
2:00 107.4 F 0.1
4:00 143.7 F 0.6
6:00 171.3 F 3.9
8:00 189.8 F 7.5
10:00 206.4 F 11.3
12:00 226.1 F 5.5
14:00 232.3 F 9.7
16:00 236.8 F 10.2
18:00 240.7 F 11.4
20:00 237.8 F 9.5
22:00 250.6 F 15.8
24:00 261.6 F 22.1
26:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 271.7 F 29.7
2:00 272.4 F 29.6
3:00 272.6 F 29.5
4:00 272.1 F 29.4

Min 271.3 F 29.2
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 227.5 F 0.0
10:00 223.9 F 0.0
15:00 179.1 F 0.0
20:00 176.1 F 0.0
25:00 175.6 F 0.0
30:00 176.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 27:08
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:05:42
CYCLE COMPLETE

04/17/2024 02:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____