
Midmark M11 - v1.0.0
Total Cycles: 3624

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2023 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.4 F 0.0
2:00 159.7 F 0.4
4:00 168.1 F 3.7
6:00 183.0 F 7.4
8:00 204.3 F 11.9
10:00 215.1 F 5.3
12:00 227.3 F 9.9
14:00 235.7 F 9.3
16:00 238.6 F 11.0
18:00 235.2 F 8.3
20:00 247.8 F 14.4
22:00 258.5 F 20.3
24:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 29.4
2:00 272.2 F 28.9
3:00 271.2 F 28.5
4:00 272.1 F 29.4

Min 271.2 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 214.6 F 0.0
10:00 221.2 F 0.0
15:00 229.7 F 0.0
20:00 229.6 F 0.0
25:00 229.5 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:20
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:03:30
CYCLE COMPLETE

08/09/2023 11:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____