

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4308

Name: RITTER 1  
Sterilizer ID: V2178198  
02/16/2024 10:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.7 F 0.0  
2:00 163.1 F 1.7  
4:00 165.8 F 4.7  
6:00 183.1 F 8.2  
8:00 205.4 F 12.9  
10:00 215.6 F 6.2  
12:00 226.5 F 10.6  
14:00 229.9 F 7.4  
16:00 241.5 F 12.3  
18:00 239.9 F 10.6  
20:00 252.1 F 16.6  
22:00 262.8 F 22.8  
24:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.6 F 29.1  
2:00 271.4 F 28.4  
3:00 272.2 F 29.3  
4:00 272.1 F 28.9

Min 271.1 F 28.4  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.5 F 0.0  
5:00 238.7 F 0.0  
10:00 230.2 F 0.0  
15:00 226.0 F 0.0  
20:00 219.8 F 0.0  
25:00 222.3 F 0.0  
30:00 225.9 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:03:19  
CYCLE COMPLETE

02/16/2024 11:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_