

-----  
Midmark M11 - v1.0.5  
Total Cycles: 426

Name: RITTER2  
Sterilizer ID: V2313228  
10/07/2021 12:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 94.0 F 0.0  
2:00 95.6 F 0.7  
4:00 123.9 F 2.7  
6:00 155.0 F 6.1  
8:00 180.2 F 10.2  
10:00 201.2 F 5.8  
12:00 214.3 F 7.1  
14:00 226.5 F 11.4  
16:00 230.7 F 7.5  
18:00 242.2 F 12.9  
20:00 241.0 F 11.1  
22:00 252.9 F 17.2  
24:00 263.4 F 23.5  
26:00 273.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.7 F 29.3  
1:00 272.5 F 29.6  
2:00 272.8 F 29.9  
3:00 272.1 F 29.6  
4:00 272.2 F 29.2

Min 271.7 F 29.1  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.4 F 0.0  
5:00 205.4 F 0.0  
10:00 197.2 F 0.0  
15:00 197.0 F 0.0  
20:00 194.3 F 0.0  
25:00 193.1 F 0.0  
30:00 192.0 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 26:48  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 01:04:51  
CYCLE COMPLETE

10/07/2021 01:22 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02