
Midmark M11 - v1.0.0
Total Cycles: 1461

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2021 02:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.7 F 0.0
2:00 187.6 F 3.4
4:00 199.0 F 7.5
6:00 221.7 F 12.6
8:00 226.4 F 6.7
10:00 238.0 F 11.8
12:00 235.9 F 8.9
14:00 240.3 F 9.9
16:00 247.8 F 14.0
18:00 260.0 F 20.7
20:00 270.9 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.7 F 29.1
2:00 272.3 F 28.6
3:00 271.5 F 28.2
4:00 272.5 F 29.1

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 203.7 F 0.0
10:00 215.5 F 0.0
15:00 213.4 F 0.0
20:00 212.9 F 0.0
25:00 215.5 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

12/15/2021 03:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____