
Midmark M11 - v1.0.5
Total Cycles: 3525

Name: RITTER2
Sterilizer ID: V2313228
04/15/2024 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.5 F 0.0
2:00 88.0 F 0.0
4:00 77.4 F 0.0
6:00 95.7 F 0.1
8:00 131.1 F 1.2
10:00 167.3 F 4.0
12:00 199.6 F 7.3
14:00 207.2 F 11.3
16:00 222.7 F 4.4
18:00 229.8 F 8.5
20:00 240.0 F 12.7
22:00 236.2 F 9.4
24:00 238.4 F 9.8
26:00 246.7 F 13.8
28:00 258.3 F 20.2
30:00 268.9 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.8
1:00 272.0 F 29.4
2:00 272.1 F 28.9
3:00 271.6 F 29.3
4:00 272.0 F 29.2

Min 271.0 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 225.0 F 0.0
10:00 194.4 F 0.0
15:00 192.6 F 0.0
20:00 187.5 F 0.0
25:00 183.8 F 0.0
30:00 181.8 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:01
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:05
CYCLE COMPLETE

04/15/2024 10:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____