
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002626
15:25 27/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:57
STERILIZING 2:52
136.0 C 220kPa 2:52
136.3 C 225kPa 3:22
136.4 C 224kPa 3:52
136.0 C 222kPa 4:22
136.4 C 226kPa 4:52
136.2 C 222kPa 5:22
136.4 C 225kPa 5:52
136.3 C 225kPa 6:22
136.3 C 225kPa 6:22
VENTING 6:22
AIR DRYING 7:14
CYCLE COMPLETE 37:14

Digital Signature #
AC5B51EEC378723A
