
Midmark M11 - v1.0.0
Total Cycles: 609

Name: MIDMARK
Sterilizer ID: V2178198
01/18/2021 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.9 F 0.0
2:00 133.9 F 0.9
4:00 142.7 F 3.5
6:00 172.3 F 7.8
8:00 199.3 F 13.0
10:00 217.3 F 7.2
12:00 230.8 F 12.5
14:00 234.1 F 9.5
16:00 236.6 F 8.7
18:00 247.7 F 14.3
20:00 260.2 F 21.3
22:00 271.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.5 F 28.9
2:00 272.1 F 29.2
3:00 271.9 F 28.6
4:00 272.6 F 29.4

Min 271.3 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 240.1 F 0.0
10:00 228.1 F 0.0
15:00 230.8 F 0.0
20:00 229.7 F 0.0
25:00 230.1 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:33
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

01/18/2021 12:20 PM
ACCEPTED: X
REJECTED:
Operator: U01