

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001424  
15:05 09/09/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:05  
PRESSURIZING 5:31  
STERILIZING 6:50  
136.0 C 220kPa 6:50  
136.3 C 222kPa 7:20  
136.4 C 220kPa 7:50  
136.5 C 222kPa 8:20  
136.2 C 222kPa 8:50  
135.8 C 218kPa 9:20  
136.2 C 222kPa 9:50  
136.4 C 223kPa 10:20  
VENTING 10:20  
AIR DRYING 11:20  
CYCLE COMPLETE 11:31  
-----

Digital Signature #  
82BF85A53944CD2B  
-----