
Midmark M11 - v1.0.0
Total Cycles: 2185

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2022 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.9 F 0.0
2:00 131.6 F 0.0
4:00 160.9 F 2.4
6:00 195.0 F 6.5
8:00 200.7 F 10.9
10:00 214.1 F 3.8
12:00 226.7 F 8.2
14:00 237.9 F 12.5
16:00 234.7 F 8.8
18:00 243.4 F 11.4
20:00 243.9 F 12.2
22:00 255.4 F 18.2
24:00 265.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.1
2:00 271.8 F 28.5
3:00 271.6 F 28.6
4:00 272.4 F 29.1

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 223.6 F 0.0
10:00 231.9 F 0.0
15:00 233.2 F 0.0
20:00 220.3 F 0.0
25:00 220.4 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 25:49
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

07/27/2022 12:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____