

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3035

Name: RITTER2  
Sterilizer ID: V2313228  
10/20/2023 10:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 109.0 F 0.0  
2:00 107.7 F 0.3  
4:00 123.1 F 1.5  
6:00 151.0 F 4.3  
8:00 174.9 F 7.9  
10:00 194.1 F 12.1  
12:00 218.7 F 5.5  
14:00 227.4 F 9.8  
16:00 233.7 F 9.5  
18:00 238.9 F 11.1  
20:00 235.6 F 9.0  
22:00 249.1 F 15.3  
24:00 260.5 F 21.6  
26:00 270.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 272.9 F 29.9  
2:00 271.5 F 29.3  
3:00 272.7 F 30.0  
4:00 271.9 F 29.4

Min 271.4 F 29.2  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
5:00 235.8 F 0.0  
10:00 196.5 F 0.0  
15:00 184.7 F 0.0  
20:00 181.2 F 0.0  
25:00 178.9 F 0.0  
30:00 177.6 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 27:34  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:05:50  
CYCLE COMPLETE

10/20/2023 11:14 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_