
Midmark M11 - v1.0.5
Total Cycles: 2688

Name: RITTER2
Sterilizer ID: V2313228
07/24/2023 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 189.4 F 3.4
4:00 200.9 F 8.3
6:00 217.3 F 13.2
8:00 226.7 F 8.0
10:00 238.4 F 13.1
12:00 239.2 F 11.0
14:00 238.0 F 9.9
16:00 251.7 F 16.5
18:00 263.0 F 23.1
20:00 272.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.3 F 29.3
2:00 272.0 F 29.5
3:00 272.0 F 29.0
4:00 272.2 F 29.7

Min 271.5 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 221.2 F 0.0
10:00 224.6 F 0.0
15:00 211.9 F 0.0
20:00 207.6 F 0.0
25:00 207.9 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 2:47
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:00:28
CYCLE COMPLETE

07/24/2023 02:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____