

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000234  
14:09 28/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:02  
PRESSURIZING 3:15  
STERILIZING 4:53  
136.4 C 232kPa 4:53  
136.2 C 232kPa 5:23  
136.2 C 231kPa 5:53  
136.4 C 233kPa 6:23  
136.4 C 233kPa 6:53  
136.4 C 231kPa 7:23  
136.4 C 233kPa 7:53  
136.4 C 233kPa 8:23  
136.4 C 233kPa 8:23  
VENTING 8:24  
AIR DRYING 9:24  
CYCLE COMPLETE 14:35  
-----

Digital Signature #  
E75D3DACF4E6461C  
-----