

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3729

Name: RITTER 1  
Sterilizer ID: V2178198  
09/06/2023 04:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
2:00 180.4 F 3.2  
4:00 212.0 F 8.6  
6:00 225.8 F 8.7  
8:00 235.2 F 9.9  
10:00 234.3 F 8.2  
12:00 244.6 F 12.4  
14:00 245.5 F 13.0  
16:00 258.6 F 20.0  
18:00 270.3 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.3 F 28.6  
2:00 272.1 F 28.9  
3:00 272.1 F 28.4  
4:00 272.3 F 29.0  
  
Min 271.1 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.5 F 0.0  
5:00 236.6 F 0.0  
10:00 209.5 F 0.0  
15:00 221.0 F 0.0  
20:00 225.1 F 0.0  
25:00 230.1 F 0.0  
30:00 221.8 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 18:49  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:58:08  
CYCLE COMPLETE

09/06/2023 05:35 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_