
Midmark M11 - v1.0.0
Total Cycles: 3690

Name: RITTER 1
Sterilizer ID: V2178198
08/28/2023 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.2 F 0.0
2:00 194.0 F 3.5
4:00 198.9 F 8.0
6:00 222.5 F 12.8
8:00 226.6 F 7.4
10:00 236.7 F 12.2
12:00 236.1 F 9.2
14:00 239.7 F 9.9
16:00 246.6 F 13.4
18:00 258.0 F 19.6
20:00 268.3 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.0 F 28.5
2:00 272.0 F 28.8
3:00 271.7 F 28.2
4:00 272.4 F 28.9

Min 271.1 F 28.1
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 233.0 F 0.0
10:00 228.2 F 0.0
15:00 222.3 F 0.0
20:00 211.5 F 0.0
25:00 233.1 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

08/28/2023 12:49 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____