
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000047
12:35 28/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:15
STERILIZING 4:41
136.4 C 225kPa 4:41
136.2 C 223kPa 5:11
136.4 C 226kPa 5:41
136.0 C 222kPa 6:11
136.2 C 223kPa 6:41
136.4 C 225kPa 7:11
136.5 C 225kPa 7:41
136.0 C 222kPa 8:11
136.0 C 222kPa 8:11
VENTING 8:12
AIR DRYING 9:02
CYCLE COMPLETE 22:11

Digital Signature #
B525594503CD228F
