

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4432

Name: RITTER 1  
Sterilizer ID: V2178198  
04/05/2024 02:28 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.1 F 0.0  
2:00 111.5 F 1.0  
4:00 129.6 F 3.1  
6:00 184.8 F 6.0  
8:00 182.7 F 9.6  
10:00 205.5 F 13.4  
12:00 214.0 F 5.2  
14:00 224.2 F 8.8  
16:00 235.2 F 12.9  
18:00 235.7 F 9.0  
20:00 241.5 F 11.4  
22:00 244.2 F 12.4  
24:00 255.4 F 18.2  
26:00 265.7 F 24.4  
28:00 271.8 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.4 F 28.9  
2:00 272.7 F 29.3  
3:00 271.4 F 28.9  
4:00 272.2 F 29.0

Min 271.3 F 28.7  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
5:00 237.7 F 0.0  
10:00 227.2 F 0.0  
15:00 221.5 F 0.0  
20:00 212.5 F 0.0  
25:00 214.5 F 0.0  
30:00 234.8 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 28:30  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:07:25  
CYCLE COMPLETE

04/05/2024 03:35 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_