
Midmark M11 - v1.0.0
Total Cycles: 1374

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2021 09:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.3 F 0.0
2:00 75.3 F 0.0
4:00 97.6 F 0.0
6:00 120.9 F 0.0
8:00 146.4 F 0.2
10:00 175.0 F 0.3
12:00 205.2 F 2.7
14:00 218.3 F 5.7
16:00 225.5 F 8.6
18:00 234.5 F 12.1
20:00 232.9 F 7.4
22:00 243.8 F 12.4
24:00 239.6 F 10.0
26:00 237.1 F 8.6
28:00 249.7 F 14.7
30:00 261.1 F 21.1
32:00 271.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.4 F 28.6
2:00 271.5 F 27.8
3:00 272.1 F 28.7
4:00 272.1 F 28.3

Min 271.1 F 27.8
Max 272.4 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 224.1 F 0.0
10:00 231.4 F 0.0
15:00 226.6 F 0.0
20:00 227.0 F 0.0
25:00 230.9 F 0.0
30:00 228.3 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:39
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:10:41
CYCLE COMPLETE

11/17/2021 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____