
Midmark M11 - v1.0.0
Total Cycles: 3862

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2023 01:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.2 F 0.0
2:00 163.9 F 3.1
4:00 188.8 F 7.0
6:00 212.9 F 11.9
8:00 227.9 F 6.4
10:00 231.8 F 11.4
12:00 233.0 F 7.9
14:00 246.0 F 12.7
16:00 244.6 F 12.3
18:00 256.7 F 18.7
20:00 267.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.8 F 28.7
2:00 272.0 F 29.0
3:00 272.1 F 29.0
4:00 272.0 F 29.0

Min 271.2 F 28.3
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 226.8 F 0.0
10:00 215.9 F 0.0
15:00 185.5 F 0.0
20:00 176.2 F 0.0
25:00 184.0 F 0.0
30:00 192.4 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

10/04/2023 02:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____