
Midmark M11 - v1.0.5
Total Cycles: 2789

Name: RITTER2
Sterilizer ID: V2313228
08/18/2023 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.1 F 0.0
2:00 169.3 F 0.0
4:00 197.2 F 0.0
6:00 209.3 F 0.0
8:00 222.9 F 4.3
10:00 233.4 F 8.9
12:00 240.7 F 11.8
14:00 241.5 F 11.5
16:00 239.5 F 10.5
18:00 239.0 F 10.2
20:00 253.4 F 17.4
22:00 265.2 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.7
1:00 272.7 F 29.5
2:00 271.9 F 28.7
3:00 272.5 F 29.5
4:00 272.2 F 29.0

Min 271.6 F 28.6
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 210.2 F 0.0
10:00 198.9 F 0.0
15:00 196.7 F 0.0
20:00 193.4 F 0.0
25:00 189.7 F 0.0
30:00 187.4 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:02:51
CYCLE COMPLETE

08/18/2023 12:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____