
Midmark M11 - v1.0.0
Total Cycles: 4694

Name: RITTER 1
Sterilizer ID: V2178198
07/03/2024 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 72.9 F 0.2
4:00 90.8 F 1.2
6:00 113.0 F 3.3
8:00 144.0 F 5.8
10:00 175.8 F 8.9
12:00 190.5 F 12.8
14:00 205.0 F 3.3
16:00 214.6 F 6.5
18:00 225.6 F 10.0
20:00 232.8 F 9.7
22:00 236.3 F 10.0
24:00 237.3 F 9.0
26:00 246.1 F 13.3
28:00 257.5 F 19.5
30:00 267.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.8 F 29.1
2:00 271.7 F 28.3
3:00 272.2 F 29.1
4:00 272.3 F 28.8

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 198.9 F 0.0
10:00 199.8 F 0.0
15:00 191.9 F 0.0
20:00 205.5 F 0.0
25:00 214.7 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:23
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:09:39
CYCLE COMPLETE

07/03/2024 09:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____