

-----  
Midmark M11 - v1.0.5  
Total Cycles: 653

Name: RITTER2  
Sterilizer ID: V2313228  
12/20/2021 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.3 F 0.0  
2:00 152.2 F 2.9  
4:00 175.5 F 6.7  
6:00 194.8 F 11.7  
8:00 215.0 F 5.9  
10:00 229.5 F 11.1  
12:00 230.2 F 7.9  
14:00 241.4 F 11.9  
16:00 244.3 F 13.0  
18:00 257.2 F 19.6  
20:00 268.3 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 29.2  
1:00 272.6 F 29.3  
2:00 272.3 F 29.5  
3:00 272.3 F 29.7  
4:00 272.0 F 29.5

Min 271.6 F 29.0  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
5:00 253.8 F 0.0  
10:00 194.2 F 0.0  
15:00 191.6 F 0.0  
20:00 190.6 F 0.0  
25:00 201.2 F 0.0  
30:00 208.9 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 21:05  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

12/20/2021 01:48 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02