

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1766

Name: RITTER2  
Sterilizer ID: V2313228  
11/23/2022 10:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.0 F 0.0  
2:00 171.0 F 0.3  
4:00 186.5 F 0.7  
6:00 213.0 F 3.9  
8:00 228.0 F 8.8  
10:00 237.3 F 10.3  
12:00 240.0 F 11.7  
14:00 238.7 F 10.2  
16:00 238.0 F 9.4  
18:00 252.8 F 16.8  
20:00 265.0 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.3 F 29.4  
2:00 271.9 F 28.8  
3:00 271.2 F 28.7  
4:00 272.2 F 29.4  
  
Min 271.1 F 28.4  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
5:00 216.3 F 0.0  
10:00 238.6 F 0.0  
15:00 234.7 F 0.0  
20:00 227.6 F 0.0  
25:00 221.0 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 21:51  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:05  
CYCLE COMPLETE

11/23/2022 11:02 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_