
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000105
8:55 27/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:41
PRESSURIZING 1:50
STERILIZING 3:10
136.4 C 224kPa 3:10
136.3 C 223kPa 3:40
136.0 C 221kPa 4:10
136.4 C 224kPa 4:40
136.1 C 221kPa 5:10
136.4 C 223kPa 5:40
136.4 C 223kPa 6:10
136.3 C 223kPa 6:40
136.3 C 223kPa 6:40
VENTING 6:40
AIR DRYING 7:34
CYCLE COMPLETE 7:42

Digital Signature #
6C5731107B5B206D
