
Midmark M11 - v1.0.0
Total Cycles: 4905

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2024 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 78.1 F 0.0
4:00 83.7 F 0.0
6:00 116.1 F 0.1
8:00 146.0 F 0.3
10:00 171.9 F 0.6
12:00 195.3 F 2.8
14:00 209.2 F 6.1
16:00 219.1 F 9.6
18:00 230.1 F 13.5
20:00 231.6 F 8.4
22:00 242.7 F 13.3
24:00 240.7 F 10.8
26:00 237.1 F 8.9
28:00 251.4 F 15.7
30:00 262.9 F 22.3
32:00 272.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.7 F 28.7
2:00 271.8 F 27.9
3:00 272.1 F 28.4
4:00 272.5 F 28.5

Min 271.4 F 27.7
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.1 F 0.0
5:00 189.5 F 0.0
10:00 211.3 F 0.0
15:00 214.2 F 0.0
20:00 215.4 F 0.0
25:00 219.3 F 0.0
30:00 221.3 F 0.0

Drying Complete

FILLING: 0:40
HEATING: 32:17
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:10:23
CYCLE COMPLETE

09/13/2024 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____