
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 004385
14:52 19/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:32
STERILIZING 3:21
136.0 C 220kPa 3:21
136.1 C 219kPa 3:51
136.3 C 224kPa 4:21
136.4 C 225kPa 4:51
136.4 C 223kPa 5:21
136.2 C 223kPa 5:51
136.4 C 224kPa 6:21
136.2 C 219kPa 6:51
136.2 C 219kPa 6:51
VENTING 6:51
AIR DRYING 8:46
CYCLE COMPLETE 11:06

Digital Signature #
74C7061CDF540F43
