
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002040
7:59 16/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:39
STERILIZING 3:44
136.4 C 223kPa 3:44
136.4 C 225kPa 4:14
136.4 C 223kPa 4:44
136.4 C 224kPa 5:14
136.4 C 225kPa 5:44
136.4 C 223kPa 6:14
136.2 C 221kPa 6:44
136.4 C 224kPa 7:14
136.4 C 224kPa 7:14
VENTING 7:14
AIR DRYING 8:05
CYCLE COMPLETE 38:05

Digital Signature #
888A40B3118E767E
