
Midmark M11 - v1.0.0
Total Cycles: 4767

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2024 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.4 F 0.0
2:00 72.0 F 0.0
4:00 87.5 F 0.1
6:00 121.1 F 1.2
8:00 145.7 F 3.5
10:00 171.1 F 6.2
12:00 191.3 F 9.5
14:00 204.9 F 11.2
16:00 213.5 F 4.8
18:00 222.6 F 8.3
20:00 233.3 F 12.2
22:00 232.2 F 7.9
24:00 244.0 F 13.1
26:00 241.5 F 10.9
28:00 253.6 F 17.1
30:00 264.3 F 23.4
32:00 271.5 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.2
3:00 272.7 F 28.8
4:00 271.5 F 28.0

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 197.5 F 0.0
10:00 198.0 F 0.0
15:00 209.0 F 0.0
20:00 213.3 F 0.0
25:00 222.2 F 0.0
30:00 225.8 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:04
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:16
CYCLE COMPLETE

07/26/2024 09:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____