
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005782
12:19 19/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:55
STERILIZING 2:51
136.0 C 220kPa 2:51
136.5 C 225kPa 3:21
136.3 C 223kPa 3:51
136.2 C 223kPa 4:21
136.2 C 224kPa 4:51
136.4 C 225kPa 5:21
136.4 C 225kPa 5:51
136.4 C 223kPa 6:21
136.4 C 223kPa 6:21
VENTING 6:21
AIR DRYING 7:15
CYCLE COMPLETE 7:26

Digital Signature #
9708B6CA115C836C
