
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004874
12:39 20/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:19
STERILIZING 3:02
136.0 C 220kPa 3:02
136.0 C 220kPa 3:33
136.4 C 224kPa 4:03
136.4 C 223kPa 4:33
136.4 C 224kPa 5:03
136.3 C 221kPa 5:33
136.3 C 224kPa 6:03
136.2 C 217kPa 6:33
136.2 C 217kPa 6:33
VENTING 6:33
AIR DRYING 7:27
CYCLE COMPLETE 10:15

Digital Signature #
F53734A8277FDC4F
