

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002240  
12:08 11/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:58  
STERILIZING 3:05  
136.0 C 220kPa 3:05  
135.9 C 223kPa 3:35  
136.0 C 222kPa 4:05  
136.4 C 225kPa 4:35  
136.4 C 223kPa 5:05  
136.3 C 225kPa 5:35  
136.4 C 227kPa 6:05  
136.4 C 223kPa 6:35  
136.4 C 223kPa 6:35  
VENTING 6:35  
AIR DRYING 7:29  
CYCLE COMPLETE 12:41  
-----

Digital Signature #  
DAD96AEC5818F630  
-----