
Midmark M11 - v1.0.0
Total Cycles: 4473

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2024 01:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.8 F 0.0
2:00 113.5 F 1.2
4:00 123.0 F 3.8
6:00 153.2 F 7.2
8:00 177.2 F 11.2
10:00 209.3 F 3.5
12:00 217.1 F 7.7
14:00 228.9 F 12.1
16:00 230.9 F 8.0
18:00 242.6 F 13.2
20:00 241.0 F 11.2
22:00 253.1 F 17.3
24:00 263.9 F 23.5
26:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.5 F 29.2
2:00 271.2 F 28.7
3:00 272.4 F 29.2
4:00 271.2 F 28.6

Min 271.2 F 28.6
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 216.2 F 0.0
10:00 215.5 F 0.0
15:00 212.2 F 0.0
20:00 219.2 F 0.0
25:00 231.8 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 26:12
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:04:45
CYCLE COMPLETE

04/18/2024 02:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____