
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007947
15:01 08/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:30
STERILIZING 3:33
136.0 C 220kPa 3:33
136.4 C 224kPa 4:03
136.4 C 224kPa 4:33
136.3 C 221kPa 5:03
136.4 C 225kPa 5:33
136.2 C 219kPa 6:03
136.4 C 224kPa 6:33
136.3 C 222kPa 7:03
136.3 C 222kPa 7:03
VENTING 7:03
AIR DRYING 7:58
CYCLE COMPLETE 8:01

Digital Signature #
01CE64F651214AAD
