
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002129
10:27 07/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:58
PRESSURIZING 3:59
STERILIZING 5:06
136.1 C 220kPa 5:06
136.4 C 225kPa 5:36
136.4 C 225kPa 6:06
136.4 C 225kPa 6:36
136.1 C 221kPa 7:06
136.4 C 225kPa 7:36
136.4 C 225kPa 8:06
136.4 C 225kPa 8:36
136.4 C 225kPa 8:36
VENTING 8:36
AIR DRYING 9:31
CYCLE COMPLETE 9:56

Digital Signature #
F0BD4B0B5CAE26A7
