

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1262

Name: RITTER 1  
Sterilizer ID: V2178198  
10/13/2021 12:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.8 F 0.0  
2:00 122.8 F 0.1  
4:00 140.7 F 0.4  
6:00 175.0 F 3.2  
8:00 198.5 F 6.7  
10:00 209.3 F 11.0  
12:00 223.6 F 5.3  
14:00 233.1 F 9.9  
16:00 236.6 F 9.4  
18:00 243.5 F 12.5  
20:00 241.6 F 11.0  
22:00 254.5 F 17.5  
24:00 265.6 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.8 F 29.1  
2:00 272.0 F 28.5  
3:00 271.3 F 28.2  
4:00 272.7 F 29.1

Min 271.3 F 28.0  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.6 F 0.0  
5:00 224.7 F 0.0  
10:00 223.3 F 0.0  
15:00 220.3 F 0.0  
20:00 221.1 F 0.0  
Dry Cycle Aborted

FILLING: 1:02  
HEATING: 25:48  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 22:13  
TOTAL CYCLE: 00:56:25

10/13/2021 01:46 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01