
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002597
11:42 24/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:42
STERILIZING 3:44
136.0 C 222kPa 3:44
136.2 C 223kPa 4:14
136.4 C 223kPa 4:44
136.4 C 223kPa 5:14
136.4 C 226kPa 5:44
136.3 C 225kPa 6:14
136.4 C 226kPa 6:44
136.3 C 226kPa 7:14
136.3 C 226kPa 7:14
VENTING 7:14
AIR DRYING 8:11
CYCLE COMPLETE 9:09

Digital Signature #
141526B2FA5EB455
