
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004353
8:42 14/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:11
STERILIZING 4:16
136.3 C 221kPa 4:16
136.4 C 223kPa 4:47
136.2 C 218kPa 5:17
135.9 C 220kPa 5:47
136.2 C 221kPa 6:17
136.2 C 219kPa 6:47
136.4 C 221kPa 7:17
136.1 C 221kPa 7:47
136.1 C 221kPa 7:47
VENTING 7:47
AIR DRYING 8:34
CYCLE COMPLETE 14:32

Digital Signature #
9DBF3F6D1BE30C34
