
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003625
10:05 09/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:01
STERILIZING 3:02
136.0 C 220kPa 3:02
136.3 C 222kPa 3:33
CYCLE ABORTED 3:54
STOP BUTTON PRESSED
NOT STERILE
END TIME 3:54

Digital Signature #
51DF53FC8D70A156
