

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3671

Name: RITTER2  
Sterilizer ID: V2313228  
06/05/2024 03:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.3 F 0.0  
2:00 128.7 F 1.6  
4:00 143.9 F 4.3  
6:00 169.6 F 8.1  
8:00 193.0 F 12.8  
10:00 222.3 F 6.0  
12:00 236.3 F 10.7  
14:00 229.9 F 7.4  
16:00 242.4 F 12.8  
18:00 240.7 F 11.3  
20:00 253.2 F 17.6  
22:00 264.6 F 24.2  
24:00 270.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.6  
1:00 271.9 F 29.7  
2:00 271.8 F 29.7  
3:00 271.8 F 29.7  
4:00 271.9 F 29.6

Min 271.8 F 29.4  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.6 F 0.0  
5:00 244.8 F 0.0  
10:00 184.2 F 0.0  
15:00 180.0 F 0.0  
20:00 176.9 F 0.0  
25:00 176.0 F 0.0  
30:00 175.8 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 24:39  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:02:56  
CYCLE COMPLETE

06/05/2024 04:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_