
Midmark M11 - v1.0.5
Total Cycles: 2294

Name: RITTER2
Sterilizer ID: V2313228
04/24/2023 03:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.7 F 0.0
2:00 195.1 F 5.0
4:00 211.9 F 10.4
6:00 225.6 F 6.2
8:00 234.3 F 12.5
10:00 239.3 F 10.9
12:00 239.2 F 10.3
14:00 254.1 F 17.7
16:00 266.5 F 25.3
18:00 272.9 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.9 F 29.5
1:00 272.6 F 29.4
2:00 272.6 F 29.5
3:00 272.7 F 29.6
4:00 272.5 F 29.8

Min 272.1 F 29.4
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 226.6 F 0.0
10:00 233.7 F 0.0
15:00 232.2 F 0.0
20:00 225.4 F 0.0
25:00 221.7 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 18:03
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:57:23
CYCLE COMPLETE

04/24/2023 04:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____