

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.3uS / 0.8ppm  
CYCLE NUMBER 000624  
15:27 01/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:23  
PRESSURIZING 2:30  
STERILIZING 3:35  
136.0 C 228kPa 3:35  
136.4 C 232kPa 4:05  
136.2 C 231kPa 4:35  
136.1 C 230kPa 5:05  
136.4 C 232kPa 5:35  
136.4 C 233kPa 6:05  
136.4 C 233kPa 6:35  
136.3 C 230kPa 7:05  
136.3 C 230kPa 7:05  
VENTING 7:05  
AIR DRYING 7:55  
CYCLE COMPLETE 22:55  
-----

Digital Signature #  
C1AE60F76EFDB92D  
-----