
Midmark M11 - v1.0.0
Total Cycles: 3224

Name: RITTER 1
Sterilizer ID: V2178198
05/12/2023 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.6 F 0.0
2:00 189.8 F 0.0
4:00 197.8 F 3.8
6:00 216.4 F 8.8
8:00 228.6 F 9.7
10:00 234.4 F 9.7
12:00 238.7 F 9.1
14:00 244.8 F 12.9
16:00 242.8 F 11.6
18:00 255.7 F 18.2
20:00 266.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.3 F 28.9
2:00 271.8 F 28.3
3:00 272.0 F 28.7
4:00 272.3 F 28.6

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 218.7 F 0.0
10:00 219.4 F 0.0
15:00 222.7 F 0.0
20:00 217.8 F 0.0
25:00 222.8 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

05/12/2023 11:21 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____