
Midmark M11 - v1.0.0
Total Cycles: 2276

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2022 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.8 F 0.0
2:00 167.6 F 0.3
4:00 179.7 F 1.8
6:00 199.1 F 6.1
8:00 219.4 F 10.9
10:00 224.7 F 6.1
12:00 237.2 F 11.6
14:00 235.5 F 9.0
16:00 238.7 F 9.7
18:00 247.0 F 13.9
20:00 258.9 F 20.4
22:00 269.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.3
2:00 271.8 F 28.6
3:00 271.8 F 28.9
4:00 272.5 F 29.2

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 237.6 F 0.0
10:00 234.4 F 0.0
15:00 229.7 F 0.0
20:00 229.7 F 0.0
25:00 230.4 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:59
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

08/31/2022 04:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____