
Midmark M11 - v1.0.5
Total Cycles: 230

Name: RITTER2
Sterilizer ID: V2313228
07/21/2021 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 179.1 F 2.9
4:00 177.1 F 7.1
6:00 193.9 F 12.0
8:00 210.0 F 6.5
10:00 225.2 F 11.8
12:00 234.4 F 9.6
14:00 237.3 F 9.3
16:00 248.7 F 15.0
18:00 260.2 F 21.7
20:00 271.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.8
1:00 272.4 F 29.8
2:00 271.7 F 29.9
3:00 271.9 F 30.1
4:00 272.4 F 29.8

Min 271.5 F 29.7
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 240.3 F 0.0
10:00 229.3 F 0.0
15:00 231.3 F 0.0
20:00 225.4 F 0.0
25:00 231.8 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

07/21/2021 10:40 AM
ACCEPTED: X
REJECTED:
Operator: U01