

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2342

Name: RITTER2  
Sterilizer ID: V2313228  
05/04/2023 03:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.2 F 0.0  
2:00 165.1 F 3.2  
4:00 178.8 F 7.2  
6:00 200.8 F 12.1  
8:00 220.1 F 6.9  
10:00 234.0 F 12.3  
12:00 234.3 F 9.6  
14:00 236.9 F 9.3  
16:00 248.3 F 15.0  
18:00 260.4 F 21.8  
20:00 271.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.3 F 29.5  
2:00 271.3 F 29.4  
3:00 272.1 F 29.5  
4:00 271.1 F 29.2

Min 271.1 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 212.1 F 0.0  
10:00 214.9 F 0.0  
15:00 218.2 F 0.0  
20:00 223.4 F 0.0  
25:00 225.6 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 20:34  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 00:59:22  
CYCLE COMPLETE

05/04/2023 04:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_