
Midmark M11 - v1.0.5
Total Cycles: 864

Name: RITTER2
Sterilizer ID: V2313228
03/09/2022 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 179.2 F 2.6
4:00 195.1 F 7.6
6:00 213.2 F 13.4
8:00 228.1 F 9.2
10:00 235.8 F 9.6
12:00 243.7 F 13.4
14:00 244.1 F 12.7
16:00 257.1 F 19.7
18:00 268.6 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.6 F 29.9
2:00 272.0 F 29.2
3:00 271.1 F 29.2
4:00 272.7 F 30.3

Min 271.0 F 28.9
Max 273.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 218.5 F 0.0
10:00 201.9 F 0.0
15:00 202.7 F 0.0
20:00 225.1 F 0.0
25:00 228.6 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

03/09/2022 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____