
Midmark M11 - v1.0.0
Total Cycles: 2342

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2022 09:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.5 F 0.0
2:00 187.2 F 3.6
4:00 189.6 F 8.2
6:00 209.4 F 13.5
8:00 230.4 F 8.1
10:00 241.8 F 13.3
12:00 239.6 F 10.4
14:00 235.5 F 8.3
16:00 250.3 F 15.3
18:00 262.1 F 22.1
20:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 273.1 F 29.4
2:00 272.0 F 28.7
3:00 272.2 F 29.1
4:00 272.3 F 29.1

Min 271.0 F 28.1
Max 273.2 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 227.1 F 0.0
10:00 224.3 F 0.0
15:00 220.1 F 0.0
20:00 218.1 F 0.0
25:00 226.4 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:20
CYCLE COMPLETE

09/20/2022 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____