
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000283
9:42 02/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:14
STERILIZING 4:19
136.0 C 227kPa 4:19
136.4 C 233kPa 4:49
136.2 C 230kPa 5:19
136.2 C 232kPa 5:49
136.4 C 232kPa 6:19
136.4 C 233kPa 6:49
136.4 C 232kPa 7:19
136.4 C 232kPa 7:49
136.4 C 232kPa 7:49
VENTING 7:49
AIR DRYING 8:44
CYCLE COMPLETE 23:44

Digital Signature #
31A0F6BC54CA2774
